

# Certificate of Analysis

## Description

Product Name	NextSeq <sup>™</sup> 550Dx High Output Reagent Kit v2.5 (300 cycles)	Catalog Number	20028871
Part Number	20028871	Lot Number	A159653

#### **Kit lot Contents**

Material	Box Part Number	Box Lot number	Cartridge/ Tube Part Number	Cartridge/ Tube Lot Number
NextSeq 550Dx High Output Reagent Cartridge v2 (300 cycles)	20019555	A159653-1	20005418	20454048
NextSeq 550Dx Buffer Cartridge v2 (300 cycles)	20019556	A159653-2	20005420	20454049
NextSeq 550Dx High Output Flowcell Cartridge v2.5 (300 cycles)	20026365	A159653-3	20026363	20454764
NextSeq 550Dx Accessory Box (300 cycles)	20019558	A159653-4	20018864	20454457

#### **Test Conditions**

Kitted reagents were tested on a NextSeq 550Dx sequencing system in a 2x151 cycle paired end run configuration with PhiX at a concentration which produced a cluster density of 160-230 K/mm². Flow Cells included in the kit lot were manufactured and released in accordance with production specifications.

#### **Test Results**

Metric	Specification	UOM	Result	
Sequencing output	≥ 90	Gigabases	Pass	- Trobs
Cumulative Q-score¹ ≥ 30	≥ 75	%	Pass	

<sup>&</sup>lt;sup>1</sup>Q-Scores measure the probability that a base is called incorrectly. A higher quality score indicates a smaller probability of error. A quality score of 30 represents an error rate of 1 in 1000, with a corresponding call accuracy of 99.9%.

### Certification

Quality Rev	iow				
edanty itev	IC 44				Y-1
Print Name	Tang Wei Qiang	Signature	XV	Date	05-JUN-2020

Document Number: 1000000074837 v00, Effective Date: 12-FEB-2019

Template Number: 1000000017665 Ver. 00, Effective Date: 31-JUL-2018