

# Certificate of Analysis

## Description

| Product Name | NextSeq <sup>™</sup> 550Dx High Output Reagent Kit v2.5 (75 cycles) | Catalog Number | 20028870 |  |
|--------------|---|----------------|----------|--|
| Part Number  | 20028870  | Lot Number     | A148070  |  |

#### **Kit lot Contents**

| Material   | Box Part<br>Number | Box<br>Lot number | Cartridge/<br>Tube Part<br>Number | Cartridge/<br>Tube Lot<br>Number |
|--|--------------------|-------------------|-----------------------------------|----------------------------------|
| NextSeq 550Dx High Output Reagent Cartridge v2 (75 cycles)   | 20019560           | A148070-1         | 20005419                          | 20396369                         |
| NextSeq 550Dx Buffer Cartridge v2 (75 cycles)                | 20019561           | A148070-2         | 20005420                          | 20396367                         |
| NextSeq 550Dx High Output Flowcell Cartridge v2.5 (75cycles) | 20031098           | A148070-3         | 20026363                          | 20390786                         |
| NextSeq 550Dx Accessory Box (75 cycles)                      | 20019563           | A148070-4         | 20018864                          | 20396783                         |

### **Test Conditions**

Kitted reagents were tested on a NextSeq 550Dx sequencing system in a 2x38 cycle paired end run configuration with PhiX at a concentration which produced a cluster density of 160-230 K/mm². Flow Cells included in the kit lot were manufactured and released in accordance with production specifications.

#### **Test Results**

| Metric                   | Specification | MOU       | Result |  |
|--------------------------|---------------|-----------|--------|--|
| Sequencing output        | ≥ 22.5        | Gigabases | Pass   |  |
| Cumulative Q-score¹ ≥ 30 | ≥ 80          | %         | Pass   |  |

<sup>&</sup>lt;sup>1</sup>Q-Scores measure the probability that a base is called incorrectly. A higher quality score indicates a smaller probability of error. A quality score of 30 represents an error rate of 1 in 1000, with a corresponding call accuracy of 99.9%.

#### Certification

| riiis docum | ent certifies that the p | roduct(s) described | above meet qualif | ty specifications. |             |
|-------------|--------------------------|---------------------|-------------------|--------------------|-------------|
| Quality Rev | iew                      |                     |                   |                    |             |
| Print Name  | Tang Wei Qiang           | Signature           |                   | Date               | 26-SEP-2019 |

Document Number: 1000000076133 v00, Effective Date: 12-FEB-2019

Template Number: 1000000017665 Ver. 00, Effective Date: 31-JUL-2018