

Certificate of Analysis

Description

Product Name	NextSeq [™] 550Dx High Output Reagent Kit v2.5 (75 cycles)	Catalog Number	20028870
Part Number	20028870	Lot Number	A148082

Kit lot Contents

Material	Box Part Bo Number Lot nu		Cartridge/ Tube Part Number		
NextSeq 550Dx High Output Reagent Cartridge v2 (75 cycles)	20019560	A148082-1	20005419	20424285	
NextSeq 550Dx Buffer Cartridge v2 (75 cycles)	20019561	A148082-2	20005420	20420062	
NextSeq 550Dx High Output Flowcell Cartridge v2.5 (75cycles)	20031098	A148082-3	20026363	20420049	
NextSeq 550Dx Accessory Box (75 cycles)	20019563	A148082-4	20018864	20420240	

Test Conditions

Kitted reagents were tested on a NextSeq 550Dx sequencing system in a 2x38 cycle paired end run configuration with PhiX at a concentration which produced a cluster density of 160-230 K/mm². Flow Cells included in the kit lot were manufactured and released in accordance with production specifications.

Test Results

Metric	Specification	UOM	Result	
Sequencing output	≥ 22.5	Gigabases	Pass	
Cumulative Q-score ¹ ≥ 30	≥ 80	%	Pass	

¹Q-Scores measure the probability that a base is called incorrectly. A higher quality score indicates a smaller probability of error. A quality score of 30 represents an error rate of 1 in 1000, with a corresponding call accuracy of 99.9%.

Certification

This document certifies that the product(s) described above meet quality specifications.					
Quality Rev	iew				
Print Name	Tang Wei Qiang	Signature	De	Date	30-JAN-2020

Document Number: 1000000076133 v00, Effective Date: 12-FEB-2019

Template Number: 1000000017665 Ver. 00, Effective Date: 31-JUL-2018