

# Certificate of Analysis

# **Description**

Product Name	NextSeq™ 550Dx High Output Reagent Kit v2.5 (75 cycles)	Catalog Number	20028870
Part Number	20028870	Lot Number	A158821

## **Kit lot Contents**

Material	Box Part Number	Box Lot number	Cartridge/ Tube Part Number	Cartridge/ Tube Lot Number
NextSeq 550Dx High Output Reagent Cartridge v2 (75 cycles)	20019560	A158821-1	20005419	20454044
NextSeq 550Dx Buffer Cartridge v2 (75 cycles)	20019561	A158821-2	20005420	20454045
NextSeq 550Dx High Output Flowcell Cartridge v2.5 (75cycles)	20031098	A158821-3	20026363	20454763
NextSeq 550Dx Accessory Box (75 cycles)	20019563	A158821-4	20018864	20447578

## **Test Conditions**

Kitted reagents were tested on a NextSeq 550Dx sequencing system in a 2x38 cycle paired end run configuration with PhiX at a concentration which produced a cluster density of 160-230 K/mm². Flow Cells included in the kit lot were manufactured and released in accordance with production specifications.

#### **Test Results**

Metric	Specification UOM Result		Result
Sequencing output	≥ 22.5	Gigabases	Pass
Cumulative Q-score <sup>1</sup> ≥ 30	≥ 80	%	Pass

<sup>&</sup>lt;sup>1</sup>Q-Scores measure the probability that a base is called incorrectly. A higher quality score indicates a smaller probability of error. A quality score of 30 represents an error rate of 1 in 1000, with a corresponding call accuracy of 99.9%.

#### Certification

This document certifies that the product(s) described above meet quality specifications.						
Quality Rev	iew					
Print Name	Ong Hong Lee	Signature	Jz.	Date	01-JUN-2020	

Document Number: 1000000076133 v00, Effective Date: 12-FEB-2019

Template Number: 1000000017665 Ver. 00, Effective Date: 31-JUL-2018