

# Certificate of Analysis

## **Description**

Product Name	NextSeq <sup>™</sup> 550Dx High Output Reagent Kit v2.5 (75 cycles)			
Catalog Number	20028870	Part Number	20028870	
Lot Number	A160219	Kit Expiration Date	22-FEB-2022	

## **Kit lot Contents**

## Box 1 of 4

Material	Part Number Lot number	
NextSeq 550Dx High Output Reagent Cartridge v2 (75 cycles)- Box	20019560	A160219-1
NextSeq 550Dx High Output Reagent Cartridge v2 (75 cycles)-Component	20005419	20488184

## Box 2 of 4

Material	Part Number	Lot number	
NextSeq 550Dx Buffer Cartridge v2 (75 cycles)- Box	20019561	A160219-2	
NextSeq 550Dx Buffer Cartridge v2 (75 cycles)- Component	20005420	20476699	

## Box 3 of 4

Material	Part Number	Lot number
NextSeq 550Dx High Output Flow Cell Cartridge v2.5 (75cycles)- Box	20031098	A160219-3

NOTE: Flow cells are individually serialized and do not appear in this list above.

## Box 4 of 4

Material	Part Number	Lot number	
NextSeq 550Dx Accessory Box (75 cycles)- Box	20019563	A160219-4	
NextSeq 550Dx Accessory Box (75 cycles)- Component	20018864	20471962	

Document Number: 1000000076133 v01, Effective Date: 12-JUN-2020

Template Number: 1000000017665 Ver. 00, Effective Date: 31-JUL-2018



## **Test Conditions**

Kitted reagents were tested on a NextSeq 550Dx sequencing system in a 2x38 cycle paired end run configuration with PhiX at a concentration which produced a cluster density of 160-230 K/mm². Flow Cells included in the kit lot were manufactured and released in accordance with production specifications.

## **Test Results**

Metric	Specification	UOM	Result	
Sequencing output	≥ 22.5	Gigabases	Pass	
Cumulative Q-score¹ ≥ 30	≥ 80	%	Pass	

<sup>&</sup>lt;sup>1</sup>Q-Scores measure the probability that a base is called incorrectly. A higher quality score indicates a smaller probability of error. A quality score of 30 represents an error rate of 1 in 1000, with a corresponding call accuracy of 99.9%.

#### Certification

This document certifies that the product(s) described above meet quality specifications.					
Quality Rev	iew				
Print Name	Tang Wei Qiang	Signature	Tang wer Osong &	Date	23-OCT-2020

Document Number: 1000000076133 v01, Effective Date: 12-JUN-2020

Template Number: 1000000017665 Ver. 00, Effective Date: 31-JUL-2018